

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
 Data File : VX021115.D
 Acq On : 09 Mar 2021 17:34
 Operator : JC/MD
 Sample : M1596-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-140N-20210305

Quant Time: Mar 10 00:33:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	207216	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.835	114	382609	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	383673	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	176007	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	140808	50.31	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.62%	
35) Dibromofluoromethane	5.464	113	120908	52.01	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.02%	
50) Toluene-d8	8.694	98	485867	53.60	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.20%	
62) 4-Bromofluorobenzene	11.118	95	183408	54.72	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 109.44%	
Target Compounds						
						Qvalue
16) Acetone	2.423	43	4637	4.02	ug/l	94
27) cis-1,2-Dichloroethene	4.564	96	6389	2.13	ug/l	89
44) Trichloroethene	7.188	130	13541	4.72	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
Data File : VX021115.D
Acq On : 09 Mar 2021 17:34
Operator : JC/MD
Sample : M1596-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-140N-20210305

Quant Time: Mar 10 00:33:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

