

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
 Data File : VX021116.D
 Acq On : 09 Mar 2021 17:57
 Operator : JC/MD
 Sample : M1596-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-140B-20210305

Quant Time: Mar 10 00:33:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	200463	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	368275	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	355946	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	163729	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	134995	49.86	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.72%
35) Dibromofluoromethane	5.464	113	114769	51.29	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.58%
50) Toluene-d8	8.694	98	458571	52.56	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.12%
62) 4-Bromofluorobenzene	11.118	95	176005	54.56	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	109.12%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.358	96	3567	1.59	ug/l	90
16) Acetone	2.417	43	5550	4.97	ug/l	100
21) trans-1,2-Dichloroethene	3.146	96	3397	1.35	ug/l	95
27) cis-1,2-Dichloroethene	4.564	96	78520	27.04	ug/l	89
44) Trichloroethene	7.188	130	318451	115.44	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
Data File : VX021116.D
Acq On : 09 Mar 2021 17:57
Operator : JC/MD
Sample : M1596-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-140B-20210305

Quant Time: Mar 10 00:33:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

