

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
 Data File : VX021117.D
 Acq On : 09 Mar 2021 18:20
 Operator : JC/MD
 Sample : M1596-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ROOF-DRAIN-20210305

Quant Time: Mar 10 00:34:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	193078	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	367786	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	368114	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	166066	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	136184	52.22	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.44%
35) Dibromofluoromethane	5.470	113	117879	52.75	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.50%
50) Toluene-d8	8.694	98	463107	53.15	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.30%
62) 4-Bromofluorobenzene	11.118	95	179292	55.65	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.30%
Target Compounds						
					Qvalue	
16) Acetone	2.416	43	4273	3.98	ug/l	97
27) cis-1,2-Dichloroethene	4.564	96	6289	2.25	ug/l	94
44) Trichloroethene	7.194	130	14207	5.16	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
Data File : VX021117.D
Acq On : 09 Mar 2021 18:20
Operator : JC/MD
Sample : M1596-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ROOF-DRAIN-20210305

Quant Time: Mar 10 00:34:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

