

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
 Data File : VX021118.D
 Acq On : 09 Mar 2021 18:44
 Operator : JC/MD
 Sample : M1596-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SPRING-20210305

Quant Time: Mar 10 00:34:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	194403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	362455	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	364678	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	156530	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	134146	51.09	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.18%	
35) Dibromofluoromethane	5.464	113	116117	52.73	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.46%	
50) Toluene-d8	8.694	98	459073	53.46	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.92%	
62) 4-Bromofluorobenzene	11.118	95	173853	54.76	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 109.52%	
Target Compounds						
16) Acetone	2.417	43	4208	3.89	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
Data File : VX021118.D
Acq On : 09 Mar 2021 18:44
Operator : JC/MD
Sample : M1596-20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SPRING-20210305

Quant Time: Mar 10 00:34:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

