

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
 Data File : VX021119.D
 Acq On : 09 Mar 2021 19:07
 Operator : JC/MD
 Sample : PB135055TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135055TB

Quant Time: Mar 10 00:34:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	203729	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.829	114	376245	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	371503	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	167261	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	137840	50.09	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.18%	
35) Dibromofluoromethane	5.470	113	119669	52.35	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 104.70%	
50) Toluene-d8	8.694	98	476903	53.50	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 107.00%	
62) 4-Bromofluorobenzene	11.118	95	181671	55.12	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 110.24%	
Target Compounds						Qvalue
16) Acetone	2.417	43	29350	25.88	ug/l	98
20) Methylene Chloride	2.834	84	10088	3.71	ug/l	98
43) Isopropyl Acetate	6.411	43	20555	3.55	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
Data File : VX021119.D
Acq On : 09 Mar 2021 19:07
Operator : JC/MD
Sample : PB135055TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135055TB

Quant Time: Mar 10 00:34:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

