

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
 Data File : VX021120.D
 Acq On : 09 Mar 2021 19:30
 Operator : JC/MD
 Sample : PB135055ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135055ZHE#04

Quant Time: Mar 10 00:34:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	204480	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	376516	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	384915	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	168062	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	140333	50.81	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.62%
35) Dibromofluoromethane	5.464	113	121236	53.00	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.00%
50) Toluene-d8	8.694	98	481649	54.00	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.00%
62) 4-Bromofluorobenzene	11.118	95	181726	55.10	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.20%
Target Compounds						
16) Acetone	2.417	43	28856	25.35	ug/l	100
20) Methylene Chloride	2.834	84	10169	3.72	ug/l	95
43) Isopropyl Acetate	6.411	43	20602	3.56	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
Data File : VX021120.D
Acq On : 09 Mar 2021 19:30
Operator : JC/MD
Sample : PB135055ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135055ZHE#04

Quant Time: Mar 10 00:34:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

