

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
 Data File : VX021121.D
 Acq On : 09 Mar 2021 19:53
 Operator : JC/MD
 Sample : PB135055ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB135055ZHE#05

Quant Time: Mar 10 00:35:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	197613	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.829	114	371441	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	372964	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	173203	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	134005	50.20	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.40%	
35) Dibromofluoromethane	5.470	113	116674	51.70	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.40%	
50) Toluene-d8	8.694	98	469891	53.40	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.80%	
62) 4-Bromofluorobenzene	11.118	95	183847	56.50	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 113.00%	
Target Compounds						
						Qvalue
16) Acetone	2.417	43	28956	26.33	ug/l	100
20) Methylene Chloride	2.834	84	10400	3.94	ug/l	99
43) Isopropyl Acetate	6.411	43	21356	3.74	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030921\
Data File : VX021121.D
Acq On : 09 Mar 2021 19:53
Operator : JC/MD
Sample : PB135055ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB135055ZHE#05

Quant Time: Mar 10 00:35:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

