

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030922\
 Data File : VX027356.D
 Acq On : 09 Mar 2022 17:50
 Operator : JC/MD
 Sample : N1800-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Mar 10 00:35:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	400402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	651612	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	605718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	287365	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	205539	44.019	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	88.040%		
35) Dibromofluoromethane	5.391	113	202120	48.071	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.140%		
50) Toluene-d8	8.653	98	763098	48.074	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.140%		
62) 4-Bromofluorobenzene	11.085	95	284821	47.079	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	94.160%		
Target Compounds						
67) Ethyl Benzene	10.195	91	77085	3.508	ug/l	98
95) Naphthalene	13.780	128	55728	2.574	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030922\
Data File : VX027356.D
Acq On : 09 Mar 2022 17:50
Operator : JC/MD
Sample : N1800-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Mar 10 00:35:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

