

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030922\
 Data File : VX027357.D
 Acq On : 09 Mar 2022 18:14
 Operator : JC/MD
 Sample : N1800-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 10 00:35:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	408125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	661104	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	612173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	283744	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	210063	44.137	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.280%
35) Dibromofluoromethane	5.391	113	207714	48.692	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.380%
50) Toluene-d8	8.653	98	775199	48.135	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.260%
62) 4-Bromofluorobenzene	11.085	95	283804	46.237	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.480%
Target Compounds						
67) Ethyl Benzene	10.195	91	30464	1.372	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030922\
Data File : VX027357.D
Acq On : 09 Mar 2022 18:14
Operator : JC/MD
Sample : N1800-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 10 00:35:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

