

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX030922\
 Data File : VX027349.D
 Acq On : 09 Mar 2022 11:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 10 00:33:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : Sw846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	50.000	56.643	-13.3	97	0.00
3 P	Chloromethane	50.000	44.662	10.7	82	0.00
4 C	Vinyl Chloride	50.000	49.132	1.7#	86	0.00
5 T	Bromomethane	50.000	70.413	-40.8#	115	0.00
6 T	Chloroethane	50.000	50.493	-1.0	90	0.00
7 T	Trichlorofluoromethane	50.000	57.195	-14.4	103	0.00
8 T	Diethyl Ether	50.000	52.283	-4.6	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.831	-3.7	94	0.00
10 T	Methyl Iodide	50.000	51.224	-2.4	86	0.00
11 T	Tert butyl alcohol	250.000	246.601	1.4	86	0.00
12 CM	1,1-Dichloroethene	50.000	48.758	2.5#	85	0.00
13 T	Acrolein	250.000	264.790	-5.9	100	0.00
14 T	Allyl chloride	50.000	51.503	-3.0	90	0.00
15 T	Acrylonitrile	250.000	236.504	5.4	83	0.00
16 T	Acetone	250.000	301.222	-20.5	100	0.00
17 T	Carbon Disulfide	50.000	46.727	6.5	85	0.00
18 T	Methyl Acetate	50.000	49.739	0.5	88	0.00
19 T	Methyl tert-butyl Ether	50.000	51.531	-3.1	90	0.00
20 T	Methylene Chloride	50.000	46.867	6.3	84	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.620	4.8	86	0.00
22 T	Diisopropyl ether	50.000	51.958	-3.9	90	0.00
23 T	Vinyl Acetate	250.000	275.100	-10.0	93	0.00
24 P	1,1-Dichloroethane	50.000	50.677	-1.4	89	0.00
25 T	2-Butanone	250.000	252.456	-1.0	89	0.00
26 T	2,2-Dichloropropane	50.000	56.480	-13.0	98	0.00
27 T	cis-1,2-Dichloroethene	50.000	47.457	5.1	85	0.00
28 T	Bromochloromethane	50.000	47.918	4.2	85	0.00
29 T	Tetrahydrofuran	250.000	244.671	2.1	86	0.00
30 C	Chloroform	50.000	51.902	-3.8#	91	0.00
31 T	Cyclohexane	50.000	50.487	-1.0	87	0.00
32 T	1,1,1-Trichloroethane	50.000	54.833	-9.7	95	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.956	-7.9	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	95	0.00
35 S	Dibromofluoromethane	50.000	43.663	12.7	84	0.00
36 T	1,1-Dichloropropene	50.000	46.192	7.6	90	0.00
37 T	Ethyl Acetate	50.000	45.896	8.2	90	0.00
38 T	Carbon Tetrachloride	50.000	49.875	0.3	96	0.00
39 T	Methylcyclohexane	50.000	54.114	-8.2	105	0.00
40 TM	Benzene	50.000	42.723	14.6	85	0.00
41 T	Methacrylonitrile	50.000	46.182	7.6	88	0.00
42 TM	1,2-Dichloroethane	50.000	52.120	-4.2	100	0.00
43 T	Isopropyl Acetate	50.000	47.651	4.7	92	0.00
44 TM	Trichloroethene	50.000	47.155	5.7	93	0.00
45 C	1,2-Dichloropropane	50.000	51.913	-3.8#	100	0.00
46 T	Dibromomethane	50.000	52.360	-4.7	101	0.00
47 T	Bromodichloromethane	50.000	53.948	-7.9	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX030922\
 Data File : WX027349.D
 Acq On : 09 Mar 2022 11:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 10 00:33:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : Sw846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48 T	Methyl methacrylate	50.000	52.717	-5.4	101	0.00
49 T	1,4-Dioxane	1000.000	955.567	4.4	95	0.00
50 S	Toluene-d8	50.000	49.461	1.1	96	0.00
51 T	4-Methyl-2-Pentanone	250.000	252.970	-1.2	96	0.00
52 CM	Toluene	50.000	52.207	-4.4#	100	0.00
53 T	t-1,3-Dichloropropene	50.000	54.410	-8.8	101	0.00
54 T	cis-1,3-Dichloropropene	50.000	55.157	-10.3	103	0.00
55 T	1,1,2-Trichloroethane	50.000	42.814	14.4	83	0.00
56 T	Ethyl methacrylate	50.000	44.279	11.4	83	0.00
57 T	1,3-Dichloropropane	50.000	43.632	12.7	85	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	254.778	-1.9	96	0.00
59 T	2-Hexanone	250.000	226.618	9.4	85	0.00
60 T	Dibromochloromethane	50.000	46.106	7.8	87	0.00
61 T	1,2-Dibromoethane	50.000	44.241	11.5	85	0.00
62 S	4-Bromofluorobenzene	50.000	40.095	19.8	78	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	79	0.00
64 T	Tetrachloroethene	50.000	51.450	-2.9	87	0.00
65 PM	Chlorobenzene	50.000	50.068	-0.1	83	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.072	-8.1	88	0.00
67 C	Ethyl Benzene	50.000	52.101	-4.2#	84	0.00
68 T	m/p-Xylenes	100.000	102.793	-2.8	83	0.00
69 T	o-Xylene	50.000	50.455	-0.9	81	0.00
70 T	Styrene	50.000	51.198	-2.4	81	0.00
71 P	Bromoform	50.000	51.768	-3.5	81	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	80	0.00
73 T	Isopropylbenzene	50.000	50.936	-1.9	84	0.00
74 T	N-amyl acetate	50.000	53.685	-7.4	85	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.624	4.8	80	0.00
76 T	1,2,3-Trichloropropane	50.000	50.498	-1.0	82	0.00
77 T	Bromobenzene	50.000	48.772	2.5	80	0.00
78 T	n-propylbenzene	50.000	52.447	-4.9	84	0.00
79 T	2-Chlorotoluene	50.000	49.357	1.3	82	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.514	-1.0	82	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	52.328	-4.7	82	0.00
82 T	4-Chlorotoluene	50.000	51.161	-2.3	83	0.00
83 T	tert-Butylbenzene	50.000	50.360	-0.7	82	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.879	-1.8	82	0.00
85 T	sec-Butylbenzene	50.000	51.337	-2.7	83	0.00
86 T	p-Isopropyltoluene	50.000	51.845	-3.7	83	0.00
87 T	1,3-Dichlorobenzene	50.000	48.898	2.2	81	0.00
88 T	1,4-Dichlorobenzene	50.000	48.834	2.3	83	0.00
89 T	n-Butylbenzene	50.000	54.091	-8.2	85	0.00
90 T	Hexachloroethane	50.000	53.696	-7.4	86	0.00
91 T	1,2-Dichlorobenzene	50.000	48.612	2.8	82	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	54.845	-9.7	89	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.316	-0.6	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030922\
Data File : VX027349.D
Acq On : 09 Mar 2022 11:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 10 00:33:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : Sw846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
94 T	Hexachlorobutadiene	50.000	50.043	-0.1	83	0.00
95 T	Naphthalene	50.000	50.784	-1.6	80	0.00
96 T	1,2,3-Trichlorobenzene	50.000	50.206	-0.4	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6