

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX030922\
 Data File : VX027365.D
 Acq On : 09 Mar 2022 21:21
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 10 00:37:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : Sw846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	50.000	46.114	7.8	88	0.00
3 P	Chloromethane	50.000	42.603	14.8	87	0.00
4 C	Vinyl Chloride	50.000	43.763	12.5#	86	0.00
5 T	Bromomethane	50.000	45.738	8.5	87	0.00
6 T	Chloroethane	50.000	44.666	10.7	90	0.00
7 T	Trichlorofluoromethane	50.000	44.630	10.7	90	0.00
8 T	Diethyl Ether	50.000	43.775	12.5	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	45.087	9.8	91	0.00
10 T	Methyl Iodide	50.000	50.388	-0.8	94	0.00
11 T	Tert butyl alcohol	250.000	224.505	10.2	88	0.00
12 CM	1,1-Dichloroethene	50.000	47.041	5.9#	92	0.00
13 T	Acrolein	250.000	239.835	4.1	101	0.00
14 T	Allyl chloride	50.000	43.611	12.8	85	0.00
15 T	Acrylonitrile	250.000	222.500	11.0	87	0.00
16 T	Acetone	250.000	223.494	10.6	84	0.00
17 T	Carbon Disulfide	50.000	43.099	13.8	88	0.00
18 T	Methyl Acetate	50.000	42.971	14.1	85	0.00
19 T	Methyl tert-butyl Ether	50.000	45.746	8.5	90	0.00
20 T	Methylene Chloride	50.000	44.981	10.0	91	0.00
21 T	trans-1,2-Dichloroethene	50.000	45.521	9.0	92	0.00
22 T	Diisopropyl ether	50.000	45.469	9.1	88	0.00
23 T	Vinyl Acetate	250.000	228.920	8.4	87	0.00
24 P	1,1-Dichloroethane	50.000	45.275	9.5	89	0.00
25 T	2-Butanone	250.000	215.418	13.8	85	0.00
26 T	2,2-Dichloropropane	50.000	40.754	18.5	79	0.00
27 T	cis-1,2-Dichloroethene	50.000	46.156	7.7	92	0.00
28 T	Bromochloromethane	50.000	44.440	11.1	88	0.00
29 T	Tetrahydrofuran	250.000	216.401	13.4	85	0.00
30 C	Chloroform	50.000	45.585	8.8#	90	0.00
31 T	Cyclohexane	50.000	45.838	8.3	88	0.00
32 T	1,1,1-Trichloroethane	50.000	46.590	6.8	90	0.00
33 S	1,2-Dichloroethane-d4	50.000	45.622	8.8	90	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	94	0.00
35 S	Dibromofluoromethane	50.000	49.222	1.6	94	0.00
36 T	1,1-Dichloropropene	50.000	45.868	8.3	89	0.00
37 T	Ethyl Acetate	50.000	43.975	12.0	85	0.00
38 T	Carbon Tetrachloride	50.000	47.796	4.4	91	0.00
39 T	Methylcyclohexane	50.000	46.615	6.8	90	0.00
40 TM	Benzene	50.000	46.414	7.2	91	0.00
41 T	Methacrylonitrile	50.000	45.536	8.9	86	0.00
42 TM	1,2-Dichloroethane	50.000	46.175	7.7	88	0.00
43 T	Isopropyl Acetate	50.000	45.204	9.6	86	0.00
44 TM	Trichloroethene	50.000	48.479	3.0	95	0.00
45 C	1,2-Dichloropropane	50.000	46.418	7.2#	88	0.00
46 T	Dibromomethane	50.000	47.606	4.8	91	0.00
47 T	Bromodichloromethane	50.000	48.264	3.5	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX030922\
 Data File : VX027365.D
 Acq On : 09 Mar 2022 21:21
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 10 00:37:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : Sw846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48 T	Methyl methacrylate	50.000	45.950	8.1	87	0.00
49 T	1,4-Dioxane	1000.000	977.194	2.3	96	0.00
50 S	Toluene-d8	50.000	49.367	1.3	94	0.00
51 T	4-Methyl-2-Pentanone	250.000	227.199	9.1	85	0.00
52 CM	Toluene	50.000	48.208	3.6#	92	0.00
53 T	t-1,3-Dichloropropene	50.000	48.543	2.9	89	0.00
54 T	cis-1,3-Dichloropropene	50.000	48.211	3.6	89	0.00
55 T	1,1,2-Trichloroethane	50.000	48.037	3.9	92	0.00
56 T	Ethyl methacrylate	50.000	48.961	2.1	91	0.00
57 T	1,3-Dichloropropane	50.000	47.177	5.6	91	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	233.823	6.5	87	0.00
59 T	2-Hexanone	250.000	226.432	9.4	84	0.00
60 T	Dibromochloromethane	50.000	49.773	0.5	93	0.00
61 T	1,2-Dibromoethane	50.000	49.670	0.7	94	0.00
62 S	4-Bromofluorobenzene	50.000	49.276	1.4	94	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	94	0.00
64 T	Tetrachloroethene	50.000	47.360	5.3	95	0.00
65 PM	Chlorobenzene	50.000	47.984	4.0	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.685	0.6	96	0.00
67 C	Ethyl Benzene	50.000	47.961	4.1#	92	0.00
68 T	m/p-Xylenes	100.000	97.290	2.7	94	0.00
69 T	o-Xylene	50.000	49.015	2.0	94	0.00
70 T	Styrene	50.000	49.415	1.2	93	0.00
71 P	Bromoform	50.000	50.437	-0.9	94	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	94	0.00
73 T	Isopropylbenzene	50.000	48.199	3.6	93	0.00
74 T	N-amyl acetate	50.000	45.782	8.4	85	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	45.762	8.5	90	0.00
76 T	1,2,3-Trichloropropane	50.000	45.696	8.6	87	0.00
77 T	Bromobenzene	50.000	49.415	1.2	96	0.00
78 T	n-propylbenzene	50.000	48.186	3.6	90	0.00
79 T	2-Chlorotoluene	50.000	46.694	6.6	91	0.00
80 T	1,3,5-Trimethylbenzene	50.000	47.711	4.6	91	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.825	8.3	84	0.00
82 T	4-Chlorotoluene	50.000	47.102	5.8	90	0.00
83 T	tert-Butylbenzene	50.000	48.638	2.7	93	0.00
84 T	1,2,4-Trimethylbenzene	50.000	48.042	3.9	91	0.00
85 T	sec-Butylbenzene	50.000	48.472	3.1	92	0.00
86 T	p-Isopropyltoluene	50.000	48.752	2.5	91	0.00
87 T	1,3-Dichlorobenzene	50.000	47.829	4.3	93	0.00
88 T	1,4-Dichlorobenzene	50.000	47.450	5.1	94	0.00
89 T	n-Butylbenzene	50.000	47.187	5.6	87	0.00
90 T	Hexachloroethane	50.000	49.059	1.9	93	0.00
91 T	1,2-Dichlorobenzene	50.000	47.882	4.2	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.359	3.3	92	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.158	1.7	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030922\
Data File : VX027365.D
Acq On : 09 Mar 2022 21:21
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 10 00:37:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : Sw846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
94 T	Hexachlorobutadiene	50.000	47.382	5.2	93	0.00
95 T	Naphthalene	50.000	50.510	-1.0	94	0.00
96 T	1,2,3-Trichlorobenzene	50.000	49.921	0.2	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6