

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030924\
 Data File : VX040627.D
 Acq On : 08 Mar 2024 15:06
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 08 23:53:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Mar 05 13:21:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	14088	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	69933	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	58576	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	42411	30.142	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.467%
60) 4-Bromofluorobenzene	11.079	95	32230	32.077	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	106.933%
63) Toluene-d8	8.647	98	85499	33.029	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	110.100%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030924\
Data File : VX040627.D
Acq On : 08 Mar 2024 15:06
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 08 23:53:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Mar 05 13:21:29 2024
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 4.897 min Scan# 61
Delta R.T. -0.001 min
Lab File: VX040627.D
Acq: 08 Mar 2024 15:06




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:128 Resp: 14088
Ion Ratio Lower Upper
128 100
49 172.8 0.0 444.0
130 126.5 0.0 297.8



#27
1,2-Dichloroethane-d4
Concen: 30.142 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.006 min
Lab File: VX040627.D
Acq: 08 Mar 2024 15:06

Tgt Ion: 65 Resp: 42411
Ion Ratio Lower Upper
65 100
67 49.4 42.5 63.7









#28
1,4-Difluorobenzene
Concen: 30.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. 0.006 min
Lab File: VX040627.D
Acq: 08 Mar 2024 15:06




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 69933
Ion Ratio Lower Upper
114 100
63 22.7 17.0 25.4
88 17.2 12.7 19.1



#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX040627.D
Acq: 08 Mar 2024 15:06

Tgt Ion:117 Resp: 58576
Ion Ratio Lower Upper
117 100
82 59.4 47.0 70.6
119 31.2 25.2 37.8









#60
4-Bromofluorobenzene
Concen: 32.077 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX040627.D
Acq: 08 Mar 2024 15:06

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Tgt Ion: 95 Resp: 32230
Ion Ratio Lower Upper
95 100
174 71.7 55.4 83.2
176 70.7 52.6 78.8



#63
Toluene-d8
Concen: 33.029 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX040627.D
Acq: 08 Mar 2024 15:06

Tgt Ion: 98 Resp: 85499
Ion Ratio Lower Upper
98 100
100 61.1 51.0 76.4

