

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX030924\  
 Data File : VX040628.D  
 Acq On : 08 Mar 2024 15:30  
 Operator : JC/MD  
 Sample : VSTDCCC020  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC020EC

Manual Integrations  
 APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024  
 Supervised By :Semsettin Yesilyurt 03/11/2024

Quant Time: Mar 08 23:53:48 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\624X030524W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Tue Mar 05 13:21:29 2024  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Bromochloromethane        | 4.904          | 128  | 15894      | 30.000   | ug/l   | 0.00     |
| 28) 1,4-Difluorobenzene      | 6.763          | 114  | 87434      | 30.000   | ug/l   | 0.00     |
| 57) Chlorobenzene-d5         | 10.055         | 117  | 83437      | 30.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 27) 1,2-Dichloroethane-d4    | 5.958          | 65   | 46177      | 29.090   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 110 |      | Recovery = | 96.967%  |        |          |
| 60) 4-Bromofluorobenzene     | 11.079         | 95   | 44043      | 30.774   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 63 - 112 |      | Recovery = | 102.567% |        |          |
| 63) Toluene-d8               | 8.647          | 98   | 109463     | 29.686   | ug/l   | 0.00     |
| Spiked Amount 30.000         | Range 91 - 112 |      | Recovery = | 98.967%  |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          |        |          |
|                              |                |      |            |          | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.166          | 85   | 21970      | 20.380   | ug/l   | 96       |
| 3) Chloromethane             | 1.295          | 50   | 21638      | 19.590   | ug/l   | 99       |
| 4) Vinyl Chloride            | 1.374          | 62   | 23834      | 19.966   | ug/l   | 99       |
| 5) Bromomethane              | 1.605          | 94   | 18424      | 25.271   | ug/l   | 96       |
| 6) Chloroethane              | 1.685          | 64   | 14666      | 20.912   | ug/l   | 98       |
| 7) Trichlorofluoromethane    | 1.886          | 101  | 36610      | 20.144   | ug/l   | 95       |
| 8) Diethyl Ether             | 2.136          | 74   | 12221      | 19.805   | ug/l   | 98       |
| 9) 1,1,2-Trichlorotrifluo... | 2.325          | 101  | 20435      | 20.527   | ug/l   | 94       |
| 10) 1,1-Dichloroethene       | 2.319          | 96   | 18402      | 19.415   | ug/l   | 96       |
| 11) Methyl Iodide            | 2.453          | 142  | 17658      | 15.404   | ug/l   | 91       |
| 12) Methyl Acetate           | 2.703          | 43   | 31402      | 20.113   | ug/l   | 99       |
| 13) Acrolein                 | 2.233          | 56   | 29227      | 99.298   | ug/l   | 98       |
| 14) Acrylonitrile            | 3.062          | 53   | 60200      | 97.921   | ug/l   | 98       |
| 15) Acetone                  | 2.380          | 58   | 15678      | 89.216   | ug/l   | 70       |
| 16) Carbon Disulfide         | 2.514          | 76   | 48146      | 18.483   | ug/l   | 98       |
| 17) Allyl chloride           | 2.660          | 41   | 34539      | 19.747   | ug/l   | 91       |
| 18) Methylene Chloride       | 2.788          | 84   | 20686      | 19.116   | ug/l   | 99       |
| 19) trans-1,2-Dichloroethene | 3.087          | 96   | 21119      | 20.409   | ug/l   | 94       |
| 20) Diisopropyl ether        | 3.764          | 45   | 68932      | 19.706   | ug/l # | 91       |
| 21) 1,1-Dichloroethane       | 3.605          | 63   | 40479      | 20.110   | ug/l   | 99       |
| 22) cis-1,2-Dichloroethene   | 4.489          | 96   | 23940      | 19.922   | ug/l   | 94       |
| 23) tert-Butyl Alcohol       | 2.959          | 59   | 26125m     | 89.940   | ug/l   |          |
| 24) Methyl tert-Butyl Ether  | 3.117          | 73   | 69490      | 19.866   | ug/l   | 94       |
| 25) Chloroform               | 5.093          | 83   | 43470      | 19.731   | ug/l   | 99       |
| 26) Cyclohexane              | 5.471          | 56   | 32354      | 19.916   | ug/l # | 98       |
| 29) 1,1-Dichloropropene      | 5.690          | 75   | 31457      | 20.785   | ug/l   | 98       |
| 30) 2-Butanone               | 4.556          | 43   | 89097      | 96.475   | ug/l   | 99       |
| 31) 2,2-Dichloropropane      | 4.471          | 77   | 34073      | 20.860   | ug/l   | 96       |
| 32) 1,1,1-Trichloroethane    | 5.385          | 97   | 38556      | 20.218   | ug/l   | 97       |
| 33) Carbon Tetrachloride     | 5.672          | 117  | 33192      | 20.687   | ug/l   | 96       |
| 34) Benzene                  | 6.038          | 78   | 83183      | 20.360   | ug/l   | 99       |
| 35) Methacrylonitrile        | 4.922          | 41   | 18278      | 19.279   | ug/l   | 96       |
| 36) 1,2-Dichloroethane       | 6.092          | 62   | 39202      | 20.753   | ug/l   | 96       |
| 37) Trichloroethene          | 7.129          | 130  | 22767      | 20.737   | ug/l   | 95       |
| 38) Methylcyclohexane        | 7.379          | 83   | 34985      | 21.468   | ug/l   | 98       |
| 39) 1,2-Dichloropropane      | 7.434          | 63   | 21740      | 20.899   | ug/l   | 98       |
| 40) Dibromomethane           | 7.580          | 93   | 16694      | 20.428   | ug/l   | 95       |
| 41) Bromodichloromethane     | 7.818          | 83   | 32855      | 20.209   | ug/l # | 99       |
| 42) Vinyl Acetate            | 3.721          | 43   | 319959     | 100.626  | ug/l   | 99       |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX030924\  
 Data File : VX040628.D  
 Acq On : 08 Mar 2024 15:30  
 Operator : JC/MD  
 Sample : VSTDCCC020  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC020EC

Manual Integrations  
 APPROVED


Reviewed By : Mahesh Dadoda 03/11/2024  
 Supervised By : Semsettin Yesilyurt 03/11/2024

Quant Time: Mar 08 23:53:48 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\624X030524W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Tue Mar 05 13:21:29 2024  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 43) Ethyl Acetate             | 4.715  | 43   | 33553    | 19.291  | ug/l   | 99       |
| 44) Isopropyl Acetate         | 6.342  | 43   | 55612    | 20.031  | ug/l   | 95       |
| 45) 1,4-Dioxane               | 7.659  | 88   | 9558     | 334.915 | ug/l   | 98       |
| 46) Methyl methacrylate       | 7.690  | 41   | 29082    | 21.090  | ug/l   | 89       |
| 47) n-amyl Acetate            | 10.842 | 43   | 46960    | 19.626  | ug/l   | 99       |
| 48) t-1,3-Dichloropropene     | 8.976  | 75   | 33938    | 19.751  | ug/l   | 97       |
| 49) cis-1,3-Dichloropropene   | 8.366  | 75   | 36209    | 20.686  | ug/l   | 93       |
| 50) 1,1,2-Trichloroethane     | 9.153  | 97   | 22036    | 20.300  | ug/l   | 97       |
| 51) Ethyl methacrylate        | 9.116  | 69   | 34882    | 20.411  | ug/l   | 96       |
| 52) 1,3-Dichloropropane       | 9.305  | 76   | 38262    | 20.657  | ug/l   | 99       |
| 53) Dibromochloromethane      | 9.519  | 129  | 24301    | 19.638  | ug/l   | 98       |
| 54) 1,2-Dibromoethane         | 9.610  | 107  | 23965    | 19.952  | ug/l   | 99       |
| 55) 2-Chloroethyl vinyl ether | 8.238  | 63   | 94336    | 108.397 | ug/l   | 99       |
| 56) Bromoform                 | 10.799 | 173  | 16487    | 18.784  | ug/l # | 99       |
| 58) 4-Methyl-2-Pentanone      | 8.574  | 43   | 185444   | 99.812  | ug/l   | 99       |
| 59) 2-Hexanone                | 9.427  | 43   | 147323   | 98.168  | ug/l   | 97       |
| 61) Tetrachloroethene         | 9.275  | 164  | 20825    | 21.028  | ug/l   | 95       |
| 62) Toluene                   | 8.720  | 91   | 92630    | 20.157  | ug/l   | 100      |
| 64) Chlorobenzene             | 10.080 | 112  | 56715    | 20.193  | ug/l   | 100      |
| 65) 1,1,1,2-Tetrachloroethane | 10.159 | 131  | 22101    | 20.630  | ug/l   | 97       |
| 66) Ethyl Benzene             | 10.195 | 91   | 110110   | 20.781  | ug/l   | 98       |
| 67) m/p-Xylenes               | 10.299 | 106  | 79342    | 40.551  | ug/l   | 95       |
| 68) o-Xylene                  | 10.640 | 106  | 37209    | 19.755  | ug/l   | 95       |
| 69) Styrene                   | 10.653 | 104  | 61919    | 19.566  | ug/l   | 98       |
| 70) Isopropylbenzene          | 10.964 | 105  | 103848   | 20.405  | ug/l   | 97       |
| 71) 1,1,2,2-Tetrachloroethane | 11.213 | 83   | 36032    | 20.195  | ug/l   | 99       |
| 72) 1,2,3-Trichloropropane    | 11.238 | 75   | 32313m   | 20.128  | ug/l   |          |
| 73) Bromobenzene              | 11.195 | 156  | 25374    | 21.060  | ug/l   | 94       |
| 74) n-propylbenzene           | 11.305 | 91   | 130712   | 20.795  | ug/l   | 100      |
| 75) 2-Chlorotoluene           | 11.366 | 91   | 79262    | 21.249  | ug/l   | 98       |
| 76) 1,3,5-Trimethylbenzene    | 11.451 | 105  | 93557    | 20.880  | ug/l   | 98       |
| 77) t-1,4-Dichloro-2-butene   | 11.018 | 75   | 9128     | 18.296  | ug/l   | 71       |
| 78) 4-Chlorotoluene           | 11.451 | 91   | 93075    | 20.916  | ug/l   | 96       |
| 79) tert-butylbenzene         | 11.713 | 119  | 85943    | 20.550  | ug/l   | 97       |
| 80) 1,2,4-Trimethylbenzene    | 11.750 | 105  | 90421    | 20.482  | ug/l   | 98       |
| 81) sec-Butylbenzene          | 11.890 | 105  | 114852   | 20.801  | ug/l   | 99       |
| 82) p-Isopropyltoluene        | 12.006 | 119  | 93647    | 20.721  | ug/l   | 99       |
| 83) 1,3-Dichlorobenzene       | 11.969 | 146  | 48191    | 20.556  | ug/l   | 98       |
| 84) 1,4-Dichlorobenzene       | 12.043 | 146  | 48706    | 20.316  | ug/l   | 99       |
| 85) n-Butylbenzene            | 12.335 | 91   | 88989    | 20.235  | ug/l   | 99       |
| 86) Hexachloroethane          | 12.536 | 117  | 15740    | 19.959  | ug/l   | 90       |
| 87) 1,2-Dichlorobenzene       | 12.335 | 146  | 46895    | 20.584  | ug/l   | 99       |
| 88) 1,2-Dibromo-3-Chloropr... | 12.945 | 75   | 9502     | 19.917  | ug/l   | 96       |
| 89) 1,2,4-Trichlorobenzene    | 13.585 | 180  | 28899    | 18.840  | ug/l   | 100      |
| 90) Hexachlorobutadiene       | 13.725 | 225  | 11495    | 18.533  | ug/l   | 96       |
| 91) Naphthalene               | 13.774 | 128  | 89667    | 17.921  | ug/l   | 99       |
| 92) 1,2,3-Trichlorobenzene    | 13.963 | 180  | 26821    | 17.733  | ug/l   | 99       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed





#1  
Bromochloromethane  
Concen: 30.000 ug/l  
RT: 4.904 min Scan# 61  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion:128 Resp: 15894  
Ion Ratio Lower Upper  
128 100  
49 173.4 0.0 444.0  
130 125.9 0.0 297.8




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#2  
Dichlorodifluoromethane  
Concen: 20.380 ug/l  
RT: 1.166 min Scan# 13  
Delta R.T. -0.001 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion: 85 Resp: 21970  
Ion Ratio Lower Upper  
85 100  
87 30.9 16.7 50.1





#3  
Chloromethane  
Concen: 19.590 ug/l  
RT: 1.295 min Scan# 34  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion: 50 Resp: 21638  
Ion Ratio Lower Upper  
50 100  
52 35.3 27.8 41.8




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#4  
Vinyl Chloride  
Concen: 19.966 ug/l  
RT: 1.374 min Scan# 47  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 62 Resp: 23834  
Ion Ratio Lower Upper  
62 100  
64 32.4 26.3 39.5





#5

Bromomethane

Concen: 25.271 ug/l

RT: 1.605 min Scan# 81

Delta R.T. 0.012 min

Lab File: VX040628.D

Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC020EC

Tgt Ion: 94 Resp: 18424

Ion Ratio Lower Upper

94 100





96 97.3 74.9 112.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024

Supervised By :Semsettin Yesilyurt 03/11/2024



#6

Chloroethane

Concen: 20.912 ug/l

RT: 1.685 min Scan# 98

Delta R.T. 0.019 min

Lab File: VX040628.D




Acq: 08 Mar 2024 15:30


Tgt Ion: 64 Resp: 14666

Ion Ratio Lower Upper

64 100


66 32.3 25.0 37.6





#7  
Trichlorofluoromethane  
Concen: 20.144 ug/l  
RT: 1.886 min Scan# 11  
Delta R.T. 0.012 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion:101 Resp: 36610  
Ion Ratio Lower Upper  
101 100  
103 64.9 49.1 73.7




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#8  
Diethyl Ether  
Concen: 19.805 ug/l  
RT: 2.136 min Scan# 172  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 74 Resp: 12221  
Ion Ratio Lower Upper  
74 100  
45 104.4 20.5 184.5









#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 20.527 ug/l  
RT: 2.325 min Scan# 203  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC

Tgt Ion:101 Resp: 2043  
Ion Ratio Lower Upper  
101 100  
85 46.3 0.0 95.2  
151 75.2 0.0 136.8




Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#10  
1,1-Dichloroethene  
Concen: 19.415 ug/l  
RT: 2.319 min Scan# 202  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 96 Resp: 18402  
Ion Ratio Lower Upper  
96 100  
61 177.7 138.3 207.5  
98 65.0 49.8 74.6








#11  
Methyl Iodide  
Concen: 15.404 ug/l  
RT: 2.453 min Scan# 211  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion:142 Resp: 17658  
Ion Ratio Lower Upper  
142 100  
127 55.4 23.8 71.4  
141 13.1 6.9 20.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#12  
Methyl Acetate  
Concen: 20.113 ug/l  
RT: 2.703 min Scan# 265  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion: 43 Resp: 31402  
Ion Ratio Lower Upper  
43 100  
74 22.7 17.8 26.6





#13  
Acrolein  
Concen: 99.298 ug/l  
RT: 2.233 min Scan# 188  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion: 56 Resp: 2922  
Ion Ratio Lower Upper  
56 100  
55 69.9 57.2 85.8




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#14  
Acrylonitrile  
Concen: 97.921 ug/l  
RT: 3.062 min Scan# 324  
Delta R.T. -0.001 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion: 53 Resp: 60200  
Ion Ratio Lower Upper  
53 100  
52 83.4 41.9 125.6  
51 39.1 18.3 54.9





#15  
Acetone  
Concen: 89.216 ug/l  
RT: 2.380 min Scan# 212  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC




Tgt Ion: 58 Resp: 15678  
Ion Ratio Lower Upper  
58 100  
43 373.3 249.9 374.9

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#16  
Carbon Disulfide  
Concen: 18.483 ug/l  
RT: 2.514 min Scan# 234  
Delta R.T. 0.012 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30



Tgt Ion: 76 Resp: 48146  
Ion Ratio Lower Upper  
76 100  
78 9.3 8.0 12.0





#17  
 Allyl chloride  
 Concen: 19.747 ug/l  
 RT: 2.660 min Scan# 258  
 Delta R.T. 0.000 min  
 Lab File: VX040628.D  
 Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC020EC

Tgt Ion: 41 Resp: 34539  
 Ion Ratio Lower Upper  
 41 100  
 39 80.6 36.5 109.5  
 76 33.4 19.7 59.0

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
 Supervised By :Semsettin Yesilyurt 03/11/2024



#18  
 Methylene Chloride  
 Concen: 19.116 ug/l  
 RT: 2.788 min Scan# 279  
 Delta R.T. 0.000 min  
 Lab File: VX040628.D  
 Acq: 08 Mar 2024 15:30


Tgt Ion: 84 Resp: 20686  
 Ion Ratio Lower Upper  
 84 100  
 49 131.2 103.8 155.8  
 51 40.3 31.8 47.6  
 86 65.3 50.8 76.2





#19  
trans-1,2-Dichloroethene  
Concen: 20.409 ug/l  
RT: 3.087 min Scan# 311  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion: 96 Resp: 21119  
Ion Ratio Lower Upper  
96 100  
61 140.7 117.9 176.9  
98 56.9 51.0 76.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#20  
Diisopropyl ether  
Concen: 19.706 ug/l  
RT: 3.764 min Scan# 439  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion: 45 Resp: 68932  
Ion Ratio Lower Upper  
45 100  
43 75.0 67.7 101.5  
87 24.0 21.5 32.3  
59 10.8 10.9 16.3#





#21  
1,1-Dichloroethane  
Concen: 20.110 ug/l  
RT: 3.605 min Scan# 413  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC




Tgt Ion: 63 Resp: 40479  
Ion Ratio Lower Upper  
63 100  
98 6.7 2.9 8.8  
100 3.9 2.0 5.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#22  
cis-1,2-Dichloroethene  
Concen: 19.922 ug/l  
RT: 4.489 min Scan# 558  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Tgt Ion: 96 Resp: 23940  
Ion Ratio Lower Upper  
96 100  
61 159.5 119.8 179.6  
98 62.9 50.9 76.3





#23  
tert-Butyl Alcohol  
Concen: 89.940 ug/l m  
RT: 2.959 min Scan# 30  
Delta R.T. -0.030 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion: 59 Resp: 2612  
Ion Ratio Lower Upper  
59 100  
52 0.0 0.0 0.0




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#24  
Methyl tert-Butyl Ether  
Concen: 19.866 ug/l  
RT: 3.117 min Scan# 333  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 73 Resp: 69490  
Ion Ratio Lower Upper  
73 100  
43 24.2 17.0 25.6









#25  
Chloroform  
Concen: 19.731 ug/l  
RT: 5.093 min Scan# 613  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC

Tgt Ion: 83 Resp: 43470  
Ion Ratio Lower Upper  
83 100  
85 67.8 54.6 81.8




Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#26  
Cyclohexane  
Concen: 19.916 ug/l  
RT: 5.471 min Scan# 719  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 56 Resp: 32354  
Ion Ratio Lower Upper  
56 100  
89 0.0 0.0 0.0  
84 82.4 67.2 100.8











#27  
1,2-Dichloroethane-d4  
Concen: 29.090 ug/l  
RT: 5.958 min Scan# 798  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC

Tgt Ion: 65 Resp: 4617  
Ion Ratio Lower Upper  
65 100  
67 48.9 42.5 63.7




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#28  
1,4-Difluorobenzene  
Concen: 30.000 ug/l  
RT: 6.763 min Scan# 931  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion:114 Resp: 87434  
Ion Ratio Lower Upper  
114 100  
63 22.2 17.0 25.4  
88 16.2 12.7 19.1





#29  
1,1-Dichloropropene  
Concen: 20.785 ug/l  
RT: 5.690 min Scan# 754  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC020EC

Tgt Ion: 75 Resp: 3145  
Ion Ratio Lower Upper  
75 100  
110 33.5 0.0 71.2  
77 31.2 0.0 62.6

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#30  
2-Butanone  
Concen: 96.475 ug/l  
RT: 4.556 min Scan# 569  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion: 43 Resp: 89097  
Ion Ratio Lower Upper  
43 100  
57 8.5 6.6 10.0  
72 25.4 20.6 31.0





#31  
2,2-Dichloropropane  
Concen: 20.860 ug/l  
RT: 4.471 min Scan# 51  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC




Tgt Ion: 77 Resp: 3407  
Ion Ratio Lower Upper  
77 100  
97 20.8 18.2 27.4

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#32  
1,1,1-Trichloroethane  
Concen: 20.218 ug/l  
RT: 5.385 min Scan# 705  
Delta R.T. 0.012 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Tgt Ion: 97 Resp: 38556  
Ion Ratio Lower Upper  
97 100  
99 63.6 52.6 79.0  
61 48.8 36.9 55.3





#33  
Carbon Tetrachloride  
Concen: 20.687 ug/l  
RT: 5.672 min Scan# 752  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion:117 Resp: 33192  
Ion Ratio Lower Upper  
117 100  
119 93.2 78.2 117.4  
121 29.3 24.7 37.1




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#34  
Benzene  
Concen: 20.360 ug/l  
RT: 6.038 min Scan# 812  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 78 Resp: 83183  
Ion Ratio Lower Upper  
78 100  
77 24.1 19.3 28.9  
51 18.6 14.4 21.6






#35  
Methacrylonitrile  
Concen: 19.279 ug/l  
RT: 4.922 min Scan# 61  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC020EC






Tgt Ion: 41 Resp: 18278  
Ion Ratio Lower Upper  
41 100  
39 61.4 28.2 84.7  
67 70.2 34.5 103.5  
52 34.6 16.3 48.8

Manual Integrations




APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#36  
1,2-Dichloroethane  
Concen: 20.753 ug/l  
RT: 6.092 min Scan# 821  
Delta R.T. 0.012 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion: 62 Resp: 39202  
Ion Ratio Lower Upper  
62 100  
98 7.2 7.0 10.4





#37  
Trichloroethene  
Concen: 20.737 ug/l  
RT: 7.129 min Scan# 991  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion:130 Resp: 2276  
Ion Ratio Lower Upper  
130 100  
95 101.8 85.8 128.8




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#38  
Methylcyclohexane  
Concen: 21.468 ug/l  
RT: 7.379 min Scan# 1032  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 83 Resp: 34985  
Ion Ratio Lower Upper  
83 100  
55 79.4 38.6 116.0  
98 45.8 22.3 66.8





#39

1,2-Dichloropropane

Concen: 20.899 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. 0.006 min

Lab File: VX040628.D

Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC020EC

Tgt Ion: 63 Resp: 21740

Ion Ratio Lower Upper

63 100





65 29.7 24.6 37.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024

Supervised By :Semsettin Yesilyurt 03/11/2024



#40

Dibromomethane

Concen: 20.428 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VX040628.D

Acq: 08 Mar 2024 15:30




Tgt Ion: 93 Resp: 16694


Ion Ratio Lower Upper

93 100

95 83.9 0.0 163.4


174 89.1 0.0 166.4





#41  
Bromodichloromethane  
Concen: 20.209 ug/l  
RT: 7.818 min Scan# 1104  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion: 83 Resp: 3285  
Ion Ratio Lower Upper  
83 100  
85 64.0 50.8 76.2  
127 10.6 5.9 8.9




Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#42  
Vinyl Acetate  
Concen: 100.626 ug/l  
RT: 3.721 min Scan# 432  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 43 Resp: 319959  
Ion Ratio Lower Upper  
43 100  
86 8.6 6.7 10.1








#43  
Ethyl Acetate  
Concen: 19.291 ug/l  
RT: 4.715 min Scan# 595  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion: 43 Resp: 3355  
Ion Ratio Lower Upper  
43 100  
45 14.2 11.8 17.8




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#44  
Isopropyl Acetate  
Concen: 20.031 ug/l  
RT: 6.342 min Scan# 862  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 43 Resp: 55612  
Ion Ratio Lower Upper  
43 100  
61 17.3 15.6 23.4





#45

1,4-Dioxane

Concen: 334.915 ug/l

RT: 7.659 min Scan# 1078

Delta R.T. 0.000 min

Lab File: VX040628.D

Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC020EC

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 88    | Resp: | 955   |
| Ion      | Ratio | Lower | Upper |
| 88       | 100   |       |       |
| 43       | 36.0  | 28.1  | 42.1  |
| 58       | 73.9  | 57.7  | 86.5  |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#46

Methyl methacrylate

Concen: 21.090 ug/l




RT: 7.690 min Scan# 1083


Delta R.T. -0.000 min

Lab File: VX040628.D

Acq: 08 Mar 2024 15:30


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 41    | Resp: | 29082 |
| Ion      | Ratio | Lower | Upper |
| 41       | 100   |       |       |
| 69       | 67.7  | 39.6  | 118.8 |
| 39       | 67.1  | 30.1  | 90.3  |





#47  
n-amy1 Acetate  
Concen: 19.626 ug/l  
RT: 10.842 min Scan# 1600  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion: 43 Resp: 46960  
Ion Ratio Lower Upper  
43 100  
55 23.1 18.2 27.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#48  
t-1,3-Dichloropropene  
Concen: 19.751 ug/l  
RT: 8.976 min Scan# 1294  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 75 Resp: 33938  
Ion Ratio Lower Upper  
75 100  
77 31.9 26.8 40.2





#49

cis-1,3-Dichloropropene

Concen: 20.686 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX040628.D

Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC020EC





Tgt Ion: 75 Resp: 36209

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 75  | 100   |       |       |
| 77  | 30.8  | 24.4  | 36.6  |
| 39  | 58.7  | 40.6  | 60.8  |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#50

1,1,2-Trichloroethane

Concen: 20.300 ug/l

RT: 9.153 min Scan# 1323




Delta R.T. -0.000 min


Lab File: VX040628.D

Acq: 08 Mar 2024 15:30

Tgt Ion: 97 Resp: 22036

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 97  | 100   |       |       |
| 83  | 84.5  | 69.0  | 103.6 |
| 85  | 55.6  | 46.3  | 69.5  |
| 99  | 61.0  | 50.9  | 76.3  |





#51

Ethyl methacrylate

Concen: 20.411 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX040628.D

Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC020EC

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 69    | Resp: | 3488  |
| Ion      | Ratio | Lower | Upper |
| 69       | 100   |       |       |
| 41       | 75.7  | 36.2  | 108.6 |
| 39       | 45.5  | 21.1  | 63.4  |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#52

1,3-Dichloropropane

Concen: 20.657 ug/l




RT: 9.305 min Scan# 1348


Delta R.T. 0.000 min

Lab File: VX040628.D

Acq: 08 Mar 2024 15:30

|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 76    | Resp: | 38262 |
| Ion      | Ratio | Lower | Upper |
| 76       | 100   |       |       |
| 78       | 32.0  | 0.0   | 65.0  |









#53  
Dibromochloromethane  
Concen: 19.638 ug/l  
RT: 9.519 min Scan# 1383  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC

Tgt Ion:129 Resp: 2430  
Ion Ratio Lower Upper  
129 100  
127 77.9 38.3 114.9




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#54  
1,2-Dibromoethane  
Concen: 19.952 ug/l  
RT: 9.610 min Scan# 1398  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion:107 Resp: 23965  
Ion Ratio Lower Upper  
107 100  
109 95.3 75.1 112.7









#55  
2-Chloroethyl vinyl ether  
Concen: 108.397 ug/l  
RT: 8.238 min Scan# 1173  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC

Tgt Ion: 63 Resp: 94330  
Ion Ratio Lower Upper  
63 100  
65 32.4 25.8 38.6  
106 26.9 20.6 31.0




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#56  
Bromoform  
Concen: 18.784 ug/l  
RT: 10.799 min Scan# 1593  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion:173 Resp: 16487  
Ion Ratio Lower Upper  
173 100  
175 47.4 37.6 56.4  
254 9.2 5.8 8.8#





#57  
Chlorobenzene-d5  
Concen: 30.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion: 117 Resp: 8343  
Ion Ratio Lower Upper  
117 100  
82 59.9 47.0 70.6  
119 32.0 25.2 37.8




Manual Integrations  
APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#58  
4-Methyl-2-Pentanone  
Concen: 99.812 ug/l  
RT: 8.574 min Scan# 1228  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 43 Resp: 185444  
Ion Ratio Lower Upper  
43 100  
58 37.1 30.1 45.1  
85 17.3 13.9 20.9








#59  
2-Hexanone  
Concen: 98.168 ug/l  
RT: 9.427 min Scan# 11  
Delta R.T. -0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion: 43 Resp: 14732  
Ion Ratio Lower Upper  
43 100  
58 50.3 42.7 64.1  
57 18.5 14.8 22.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#60  
4-Bromofluorobenzene  
Concen: 30.774 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion: 95 Resp: 44043  
Ion Ratio Lower Upper  
95 100  
174 73.4 55.4 83.2  
176 70.3 52.6 78.8





#61  
Tetrachloroethene  
Concen: 21.028 ug/l  
RT: 9.275 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion:164 Resp: 2082  
Ion Ratio Lower Upper  
164 100  
166 129.3 104.0 156.0  
129 92.6 82.0 123.0  
131 89.4 76.4 114.6




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#62  
Toluene  
Concen: 20.157 ug/l  
RT: 8.720 min Scan# 1252  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion: 91 Resp: 92630  
Ion Ratio Lower Upper  
91 100  
92 57.7 45.9 68.9





#63  
Toluene-d8  
Concen: 29.686 ug/l  
RT: 8.647 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC




Tgt Ion: 98 Resp: 10946  
Ion Ratio Lower Upper  
98 100  
100 63.4 51.0 76.4

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#64  
Chlorobenzene  
Concen: 20.193 ug/l  
RT: 10.080 min Scan# 1475  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Tgt Ion:112 Resp: 56715  
Ion Ratio Lower Upper  
112 100  
114 32.6 25.9 38.9





#65  
1,1,1,2-Tetrachloroethane  
Concen: 20.630 ug/l  
RT: 10.159 min Scan# 1488  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC




Tgt Ion:131 Resp: 2210  
Ion Ratio Lower Upper  
131 100  
133 93.5 0.0 191.8  
119 65.1 0.0 135.4

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#66  
Ethyl Benzene  
Concen: 20.781 ug/l  
RT: 10.195 min Scan# 1494  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30



Tgt Ion: 91 Resp: 110110  
Ion Ratio Lower Upper  
91 100  
106 29.5 24.4 36.6









#67  
m/p-Xylenes  
Concen: 40.551 ug/l  
RT: 10.299 min Scan# 1511  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC

Tgt Ion:106 Resp: 7934  
Ion Ratio Lower Upper  
106 100  
91 218.9 169.4 254.0




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#68  
o-Xylene  
Concen: 19.755 ug/l  
RT: 10.640 min Scan# 1567  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion:106 Resp: 37209  
Ion Ratio Lower Upper  
106 100  
91 231.2 111.2 333.6





#69  
Styrene  
Concen: 19.566 ug/l  
RT: 10.653 min Scan# 1569  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC




Tgt Ion:104 Resp: 61919  
Ion Ratio Lower Upper  
104 100  
78 57.6 43.7 65.5  
103 55.4 44.6 67.0

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#70  
Isopropylbenzene  
Concen: 20.405 ug/l  
RT: 10.964 min Scan# 1620  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30



Tgt Ion:105 Resp: 103848  
Ion Ratio Lower Upper  
105 100  
120 24.8 18.5 34.3









#71  
1,1,2,2-Tetrachloroethane  
Concen: 20.195 ug/l  
RT: 11.213 min Scan# 1661  
Delta R.T. -0.001 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC

Tgt Ion: 83 Resp: 3603  
Ion Ratio Lower Upper  
83 100  
131 10.0 4.6 13.8  
85 64.6 32.5 97.4




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#72  
1,2,3-Trichloropropane  
Concen: 20.128 ug/l m  
RT: 11.238 min Scan# 1665  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Tgt Ion: 75 Resp: 32313  
Ion Ratio Lower Upper  
75 100  
77 42.7 0.0 76.4





#73

Bromobenzene

Concen: 21.060 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX040628.D

Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC020EC

|           |       |       |       |
|-----------|-------|-------|-------|
| Tgt Ion:  | 156   | Resp: | 25374 |
| Ion Ratio | Lower | Upper |       |
| 156       | 100   |       |       |
| 77        | 170.8 | 0.0   | 363.2 |
| 158       | 94.6  | 0.0   | 197.8 |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#74

n-propylbenzene

Concen: 20.795 ug/l




RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min


Lab File: VX040628.D

Acq: 08 Mar 2024 15:30

|           |       |       |        |
|-----------|-------|-------|--------|
| Tgt Ion:  | 91    | Resp: | 130712 |
| Ion Ratio | Lower | Upper |        |
| 91        | 100   |       |        |
| 120       | 21.6  | 0.0   | 43.4   |











#75  
2-Chlorotoluene  
Concen: 21.249 ug/l  
RT: 11.366 min Scan# 1686  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30

Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC

Tgt Ion: 91 Resp: 7926  
Ion Ratio Lower Upper  
91 100  
126 31.2 0.0 64.2




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#76  
1,3,5-Trimethylbenzene  
Concen: 20.880 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion:105 Resp: 93557  
Ion Ratio Lower Upper  
105 100  
120 45.0 0.0 92.6





#77  
t-1,4-Dichloro-2-butene  
Concen: 18.296 ug/l  
RT: 11.018 min Scan# 1629  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion: 75 Resp: 9128  
Ion Ratio Lower Upper  
75 100  
53 131.4 46.9 140.8  
89 46.4 21.3 63.7




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#78  
4-Chlorotoluene  
Concen: 20.916 ug/l  
RT: 11.451 min Scan# 1700  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion: 91 Resp: 93075  
Ion Ratio Lower Upper  
91 100  
126 26.8 0.0 57.4





#79  
tert-butylbenzene  
Concen: 20.550 ug/l  
RT: 11.713 min Scan# 1743  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC




Tgt Ion:119 Resp: 8594  
Ion Ratio Lower Upper  
119 100  
91 65.5 0.0 125.6  
134 22.7 0.0 43.8

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#80  
1,2,4-Trimethylbenzene  
Concen: 20.482 ug/l  
RT: 11.750 min Scan# 1749  
Delta R.T. -0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Tgt Ion:105 Resp: 90421  
Ion Ratio Lower Upper  
105 100  
120 42.1 0.0 86.4





#81  
 sec-Butylbenzene  
 Concen: 20.801 ug/l  
 RT: 11.890 min Scan# 1772  
 Delta R.T. 0.000 min  
 Lab File: VX040628.D  
 Acq: 08 Mar 2024 15:30




Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VSTDCCC020EC




Tgt Ion:105 Resp: 114852  
 Ion Ratio Lower Upper  
 105 100  
 134 18.8 0.0 38.4

Manual Integrations  
 APPROVED



Reviewed By :Mahesh Dadoda 03/11/2024  
 Supervised By :Semsettin Yesilyurt 03/11/2024




#82  
 p-Isopropyltoluene  
 Concen: 20.721 ug/l  
 RT: 12.006 min Scan# 1791  
 Delta R.T. 0.000 min  
 Lab File: VX040628.D  
 Acq: 08 Mar 2024 15:30



Tgt Ion:119 Resp: 93647  
 Ion Ratio Lower Upper  
 119 100  
 134 25.6 0.0 52.6  
 91 27.1 0.0 54.2





#83

1,3-Dichlorobenzene

Concen: 20.556 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX040628.D


Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC020EC






|             |       |           |
|-------------|-------|-----------|
| Tgt Ion:146 | Resp: | 4819      |
| Ion Ratio   | Lower | Upper     |
| 146         | 100   |           |
| 111         | 44.7  | 22.3 66.9 |
| 148         | 63.1  | 32.8 98.3 |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#84

1,4-Dichlorobenzene


Concen: 20.316 ug/l

RT: 12.043 min Scan# 1797



Delta R.T. -0.000 min


Lab File: VX040628.D

Acq: 08 Mar 2024 15:30




|             |       |           |
|-------------|-------|-----------|
| Tgt Ion:146 | Resp: | 48706     |
| Ion Ratio   | Lower | Upper     |
| 146         | 100   |           |
| 111         | 41.5  | 21.4 64.2 |
| 148         | 64.1  | 31.8 95.4 |





#85  
n-Butylbenzene  
Concen: 20.235 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC



Tgt Ion: 91 Resp: 88989  
Ion Ratio Lower Upper  
91 100  
92 52.2 0.0 106.2  
134 23.7 0.0 48.8




Manual Integrations  
APPROVED


Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#86  
Hexachloroethane  
Concen: 19.959 ug/l  
RT: 12.536 min Scan# 1878  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30


Tgt Ion:117 Resp: 15740  
Ion Ratio Lower Upper  
117 100  
201 70.9 31.8 95.3





#87  
1,2-Dichlorobenzene  
Concen: 20.584 ug/l  
RT: 12.335 min Scan# 1845  
Delta R.T. 0.000 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30




Instrument :  
MSVOA\_X  
ClientSampleId :  
VSTDCCC020EC




Tgt Ion:146 Resp: 4689  
Ion Ratio Lower Upper  
146 100  
111 45.6 22.4 67.2  
148 63.6 31.9 95.7

Manual Integrations  
APPROVED



Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024




#88  
1,2-Dibromo-3-Chloropropane  
Concen: 19.917 ug/l  
RT: 12.945 min Scan# 1945  
Delta R.T. 0.006 min  
Lab File: VX040628.D  
Acq: 08 Mar 2024 15:30



Tgt Ion: 75 Resp: 9502  
Ion Ratio Lower Upper  
75 100  
155 69.3 58.2 87.4  
157 92.1 76.5 114.7






#89  
 1,2,4-Trichlorobenzene  
 Concen: 18.840 ug/l  
 RT: 13.585 min Scan# 2050  
 Delta R.T. -0.000 min  
 Lab File: VX040628.D  
 Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :

VSTDCCC020EC






Tgt Ion:180 Resp: 28895  
 Ion Ratio Lower Upper  
 180 100  
 182 93.7 74.5 111.7  
 145 33.4 26.6 40.0


Manual Integrations

APPROVED



Reviewed By :Mahesh Dadoda 03/11/2024  
 Supervised By :Semsettin Yesilyurt 03/11/2024




#90  
 Hexachlorobutadiene  
 Concen: 18.533 ug/l  
 RT: 13.725 min Scan# 2073  
 Delta R.T. 0.000 min  
 Lab File: VX040628.D  
 Acq: 08 Mar 2024 15:30



Tgt Ion:225 Resp: 11495  
 Ion Ratio Lower Upper  
 225 100  
 223 65.8 0.0 124.4  
 227 64.1 0.0 124.4







#91

Naphthalene

Concen: 17.921 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX040628.D

Acq: 08 Mar 2024 15:30

Instrument :

MSVOA\_X

ClientSampleId :





VSTDCCC020EC

|             |       |           |
|-------------|-------|-----------|
| Tgt Ion:128 | Resp: | 8966      |
| Ion Ratio   | Lower | Upper     |
| 128         | 100   |           |
| 127         | 13.4  | 10.2 15.2 |
| 129         | 10.9  | 8.6 12.8  |

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 03/11/2024  
Supervised By :Semsettin Yesilyurt 03/11/2024



#92

1,2,3-Trichlorobenzene

Concen: 17.733 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. 0.006 min

Lab File: VX040628.D

Acq: 08 Mar 2024 15:30

|             |       |           |
|-------------|-------|-----------|
| Tgt Ion:180 | Resp: | 26821     |
| Ion Ratio   | Lower | Upper     |
| 180         | 100   |           |
| 182         | 96.5  | 0.0 191.6 |
| 145         | 35.2  | 0.0 69.6  |

