

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031020\
 Data File : VX015157.D
 Acq On : 10 Mar 2020 09:45
 Operator : JC/SP
 Sample : VX0310WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0310WBL01

Quant Time: Mar 11 07:38:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	326439	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	519195	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	490545	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	235641	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	182711	47.18	ug/l	0.00
Spiked Amount			Recovery	=	94.36%	
35) Dibromofluoromethane	5.48	113	159138	48.59	ug/l	0.00
Spiked Amount			Recovery	=	97.18%	
50) Toluene-d8	8.71	98	625646	50.22	ug/l	0.00
Spiked Amount			Recovery	=	100.44%	
62) 4-Bromofluorobenzene	11.13	95	221666	49.88	ug/l	0.00
Spiked Amount			Recovery	=	99.76%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031020\
Data File : VX015157.D
Acq On : 10 Mar 2020 09:45
Operator : JC/SP
Sample : VX0310WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0310WBL01

Quant Time: Mar 11 07:38:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

