

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031020\
 Data File : VX015161.D
 Acq On : 10 Mar 2020 11:33
 Operator : JC/SP
 Sample : L1842-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D5-20200304

Quant Time: Mar 11 07:53:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	333743	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	560409	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	537278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	266555	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	205625	51.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.86%	
35) Dibromofluoromethane	5.49	113	172616	48.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.64%	
50) Toluene-d8	8.71	98	687627	51.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.26%	
62) 4-Bromofluorobenzene	11.14	95	248125	51.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.46%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.19	85	4558	1.560	ug/l	99
9) 1,1,2-Trichlorotrifluoroet	2.38	101	29738	9.463	ug/l	96
12) 1,1-Dichloroethene	2.37	96	6972	2.150	ug/l	94
27) cis-1,2-Dichloroethene	4.59	96	6452	1.661	ug/l	75
38) Carbon Tetrachloride	5.78	117	3324	0.680	ug/l	90
44) Trichloroethene	7.21	130	384840	89.903	ug/l	98
55) 1,1,2-Trichloroethane	9.21	97	2108	0.545	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031020\
Data File : VX015161.D
Acq On : 10 Mar 2020 11:33
Operator : JC/SP
Sample : L1842-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D5-20200304

Quant Time: Mar 11 07:53:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

