

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031020\
 Data File : VX015165.D
 Acq On : 10 Mar 2020 13:03
 Operator : JC/SP
 Sample : L1842-21
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D2-20200305

Manual Integrations
 APPROVED

MMDadoda
 3/11/2020 3:21:20 PM

Quant Time: Mar 11 12:36:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	304616	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	500260	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	496754	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	253700	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	173056	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
35) Dibromofluoromethane	5.49	113	150890	47.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.62%	
50) Toluene-d8	8.71	98	614200	51.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	11.13	95	227173	53.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.10%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	75575	26.349	ug/l	98
16) Acetone	2.44	43	6329m	4.324	ug/l	
27) cis-1,2-Dichloroethene	4.58	96	7246	2.043	ug/l	89
44) Trichloroethene	7.21	130	473943	124.031	ug/l	97
64) Tetrachloroethene	9.33	164	62033	15.103	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031020\
Data File : VX015165.D
Acq On : 10 Mar 2020 13:03
Operator : JC/SP
Sample : L1842-21
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D2-20200305

Manual Integrations
APPROVED

MMDadoda
3/11/2020 3:21:20 PM

Quant Time: Mar 11 12:36:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

