

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031020\  
 Data File : VX015172.D  
 Acq On : 10 Mar 2020 15:42  
 Operator : JC/SP  
 Sample : L1842-29DL 5X  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE125D3-20200306DL

Quant Time: Mar 11 08:44:50 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X030420W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 04 14:19:10 2020  
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	346124	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	569753	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	559189	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	281590	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	215309	52.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.86%	
35) Dibromofluoromethane	5.49	113	178188	49.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.14%	
50) Toluene-d8	8.71	98	703846	51.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.96%	
62) 4-Bromofluorobenzene	11.14	95	254352	52.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.30%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	34012	10.436	ug/l	97
44) Trichloroethene	7.21	130	150967	34.689	ug/l	96
64) Tetrachloroethene	9.33	164	5389	1.166	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031020\  
 Data File : VX015172.D  
 Acq On : 10 Mar 2020 15:42  
 Operator : JC/SP  
 Sample : L1842-29DL 5X  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 Client Sampled :  
 RE125D3-20200306DL

Quant Time: Mar 11 08:44:50 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X030420W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Mar 04 14:19:10 2020  
 Response via : Initial Calibration

