

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031020\
 Data File : VX015178.D
 Acq On : 10 Mar 2020 17:57
 Operator : JC/SP
 Sample : L1842-25DL 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D2-20200305DL

Quant Time: Mar 11 09:23:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	306337	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	498367	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	534141	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	271622	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	179525	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
35) Dibromofluoromethane	5.49	113	152992	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
50) Toluene-d8	8.71	98	641788	53.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.32%	
62) 4-Bromofluorobenzene	11.14	95	242321	56.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.62%	
Target Compounds						
44) Trichloroethene	7.21	130	141456	37.160	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031020\
Data File : VX015178.D
Acq On : 10 Mar 2020 17:57
Operator : JC/SP
Sample : L1842-25DL 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D2-20200305DL

Quant Time: Mar 11 09:23:56 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

