

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX031020\
 Data File : VX015179.D
 Acq On : 10 Mar 2020 18:20
 Operator : JC/SP
 Sample : L1842-13DL 50X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP01-20200304DL

Quant Time: Mar 11 09:25:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	326744	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	538894	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	522018	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	263285	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	203532	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
35) Dibromofluoromethane	5.49	113	170432	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
50) Toluene-d8	8.71	98	656524	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
62) 4-Bromofluorobenzene	11.14	95	237410	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
Target Compounds						
44) Trichloroethene	7.21	130	203445	49.425	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX031020\
Data File : VX015179.D
Acq On : 10 Mar 2020 18:20
Operator : JC/SP
Sample : L1842-13DL 50X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP01-20200304DL

Quant Time: Mar 11 09:25:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

