

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031021\
 Data File : VX021131.D
 Acq On : 10 Mar 2021 13:11
 Operator : JC/MD
 Sample : M1628-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW6-EB-20210309

Quant Time: Mar 11 02:10:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	204473	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.835	114	388499	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	387170	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	167431	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	141880	51.37	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.74%	
35) Dibromofluoromethane	5.464	113	121680	51.55	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.10%	
50) Toluene-d8	8.694	98	491615	53.42	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.84%	
62) 4-Bromofluorobenzene	11.118	95	185076	54.38	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 108.76%	
Target Compounds						Qvalue
11) Tert butyl alcohol	3.011	59	4425	7.19	ug/l	100
16) Acetone	2.423	43	14380	12.64	ug/l	98
25) 2-Butanone	4.647	43	2427	1.40	ug/l	92
29) Tetrahydrofuran	5.105	42	1480	1.33	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031021\
Data File : VX021131.D
Acq On : 10 Mar 2021 13:11
Operator : JC/MD
Sample : M1628-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW6-EB-20210309

Quant Time: Mar 11 02:10:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

