

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031021\
 Data File : VX021132.D
 Acq On : 10 Mar 2021 13:31
 Operator : JC/MD
 Sample : M1596-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB01-20210305

Quant Time: Mar 11 02:11:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	205496	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.835	114	376246	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	372275	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	160536	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.035	65	134726	48.54	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.08%
35) Dibromofluoromethane	5.470	113	117008	51.18	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.36%
50) Toluene-d8	8.694	98	468972	52.61	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.22%
62) 4-Bromofluorobenzene	11.118	95	175454	53.24	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.48%
Target Compounds						
11) Tert butyl alcohol	3.005	59	2175	3.52	ug/l	Qvalue # 73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031021\
Data File : VX021132.D
Acq On : 10 Mar 2021 13:31
Operator : JC/MD
Sample : M1596-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB01-20210305

Quant Time: Mar 11 02:11:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

