

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031022\
 Data File : VX027368.D
 Acq On : 10 Mar 2022 11:23
 Operator : JC/MD
 Sample : VX0310MBL01
 Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 VX0310MBL01

Quant Time: Mar 11 00:43:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	411213	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	661361	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	617752	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	284819	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	209428	43.673	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.340%
35) Dibromofluoromethane	5.385	113	210209	49.258	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.520%
50) Toluene-d8	8.653	98	779732	48.397	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.800%
62) 4-Bromofluorobenzene	11.079	95	281545	45.851	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031022\
 Data File : VX027368.D
 Acq On : 10 Mar 2022 11:23
 Operator : JC/MD
 Sample : VX0310MBL01
 Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0310MBL01

Quant Time: Mar 11 00:43:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

