

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX031022\  
 Data File : VX027369.D  
 Acq On : 10 Mar 2022 11:46  
 Operator : JC/MD  
 Sample : VX0310WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0310WBL01

Quant Time: Mar 11 00:43:54 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X030822W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Mar 08 15:54:34 2022  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 5.556  | 168  | 376403   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.763  | 114  | 610527   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.055 | 117  | 567257   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.024 | 152  | 253333   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |         |
|-----------------------------|--------|-------|----------|----------|------|---------|
| System Monitoring Compounds |        |       |          |          |      |         |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 198400   | 45.199   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =    | 90.400% |
| 35) Dibromofluoromethane    | 5.391  | 113   | 195220   | 49.555   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =    | 99.100% |
| 50) Toluene-d8              | 8.653  | 98    | 722613   | 48.587   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =    | 97.180% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 256784   | 45.301   | ug/l | 0.00    |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =    | 90.600% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX031022\  
Data File : VX027369.D  
Acq On : 10 Mar 2022 11:46  
Operator : JC/MD  
Sample : VX0310WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0310WBL01

Quant Time: Mar 11 00:43:54 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X030822W.M  
Quant Title : SW846 8260  
QLast Update : Tue Mar 08 15:54:34 2022  
Response via : Initial Calibration

