

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031022\
 Data File : VX027372.D
 Acq On : 10 Mar 2022 13:10
 Operator : JC/MD
 Sample : N1815-02DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PW-1DL

Quant Time: Mar 11 00:45:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	402914	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	638605	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	591402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	276503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	199067	42.367	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	84.740%		
35) Dibromofluoromethane	5.391	113	199779	48.482	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.960%		
50) Toluene-d8	8.653	98	755396	48.558	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.120%		
62) 4-Bromofluorobenzene	11.079	95	273551	46.137	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	92.280%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	6520	0.700	ug/l	95
25) 2-Butanone	4.580	43	2512	0.776	ug/l	98
31) Cyclohexane	5.471	56	52296	7.957	ug/l	92
39) Methylcyclohexane	7.385	83	68290	9.453	ug/l	97
40) Benzene	6.044	78	11568	0.664	ug/l	91
52) Toluene	8.720	92	4092	0.368	ug/l	100
67) Ethyl Benzene	10.195	91	4633	0.216	ug/l	94
68) m/p-Xylenes	10.305	106	7111	0.845	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	287548	16.585	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	187679	10.915	ug/l	98
86) p-Isopropyltoluene	12.012	119	16016	0.905	ug/l	93
94) Hexachlorobutadiene	13.725	225	833	0.306	ug/l	98
95) Naphthalene	13.780	128	575273	27.618	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031022\
Data File : VX027372.D
Acq On : 10 Mar 2022 13:10
Operator : JC/MD
Sample : N1815-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PW-1DL

Quant Time: Mar 11 00:45:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

