

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031022\
 Data File : VX027374.D
 Acq On : 10 Mar 2022 13:57
 Operator : JC/MD
 Sample : N1797-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-MW161S1-GW-20220303

Quant Time: Mar 11 00:45:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	379237	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	612701	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	578086	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	268330	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	195738	44.260	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	88.520%		
35) Dibromofluoromethane	5.391	113	196740	49.763	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.520%		
50) Toluene-d8	8.653	98	720576	48.278	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.560%		
62) 4-Bromofluorobenzene	11.085	95	264702	46.532	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.060%		
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.319	96	2351	0.639	ug/l	82
19) Methyl tert-butyl Ether	3.117	73	5267	0.432	ug/l	99
24) 1,1-Dichloroethane	3.617	63	15633	2.233	ug/l	97
30) Chloroform	5.099	83	2282	0.309	ug/l	95
44) Trichloroethene	7.129	130	4486	0.977	ug/l	97
64) Tetrachloroethene	9.275	164	1864	0.460	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031022\
Data File : VX027374.D
Acq On : 10 Mar 2022 13:57
Operator : JC/MD
Sample : N1797-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-MW161S1-GW-20220303

Quant Time: Mar 11 00:45:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

