

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031022\
 Data File : VX027378.D
 Acq On : 10 Mar 2022 15:55
 Operator : JC/MD
 Sample : N1864-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Mar 11 00:46:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 15:54:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	386895	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	618619	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	584512	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	265777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	198483	43.992	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.980%
35) Dibromofluoromethane	5.391	113	196868	49.319	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.640%
50) Toluene-d8	8.653	98	734599	48.746	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.500%
62) 4-Bromofluorobenzene	11.085	95	268915	46.820	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.640%
Target Compounds						
96) 1,2,3-Trichlorobenzene	13.963	180	2387	0.401	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031022\
Data File : VX027378.D
Acq On : 10 Mar 2022 15:55
Operator : JC/MD
Sample : N1864-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Mar 11 00:46:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X030822W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 15:54:34 2022
Response via : Initial Calibration

