

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
 Data File : VX045189.D
 Acq On : 10 Mar 2025 12:33
 Operator : JC/MD
 Sample : Q1500-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D2-20250304

Quant Time: Mar 11 03:53:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	76078	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	151155	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	138329	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57557	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65001	53.714	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.420%	
35) Dibromofluoromethane	5.385	113	51885	51.334	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.660%	
50) Toluene-d8	8.647	98	188382	51.411	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.820%	
62) 4-Bromofluorobenzene	11.079	95	63599	52.378	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.760%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	821	0.857	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	15311	17.316	ug/l	99
16) Acetone	2.380	43	832	1.482	ug/l	94
27) cis-1,2-Dichloroethene	4.483	96	2700	2.368	ug/l	90
38) Carbon Tetrachloride	5.666	117	1023	0.727	ug/l	89
44) Trichloroethene	7.123	130	110033	107.060	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
Data File : VX045189.D
Acq On : 10 Mar 2025 12:33
Operator : JC/MD
Sample : Q1500-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D2-20250304

Quant Time: Mar 11 03:53:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

