

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
 Data File : VX045190.D
 Acq On : 10 Mar 2025 12:57
 Operator : JC/MD
 Sample : Q1500-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW178S-20250305

Quant Time: Mar 11 03:54:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	75866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	150795	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	139303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57584	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63142	52.324	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.640%	
35) Dibromofluoromethane	5.379	113	51853	51.425	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.840%	
50) Toluene-d8	8.647	98	187213	51.214	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.420%	
62) 4-Bromofluorobenzene	11.079	95	63985	52.822	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.640%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1033	1.845	ug/l #	84
19) Methyl tert-butyl Ether	3.117	73	6392	2.044	ug/l	93
44) Trichloroethene	7.135	130	540	0.527	ug/l	81
64) Tetrachloroethene	9.275	164	334	0.374	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
Data File : VX045190.D
Acq On : 10 Mar 2025 12:57
Operator : JC/MD
Sample : Q1500-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW178S-20250305

Quant Time: Mar 11 03:54:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

