

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
 Data File : VX045191.D
 Acq On : 10 Mar 2025 13:20
 Operator : JC/MD
 Sample : Q1500-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW178I-20250305

Quant Time: Mar 11 03:54:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	73273	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144084	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	134557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57079	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60998	52.336	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125		Recovery	= 104.680%	
35) Dibromofluoromethane	5.379	113	50130	52.032	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124		Recovery	= 104.060%	
50) Toluene-d8	8.647	98	182875	52.357	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113		Recovery	= 104.720%	
62) 4-Bromofluorobenzene	11.079	95	61971	53.542	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121		Recovery	= 107.080%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1063	1.966	ug/l #	81
19) Methyl tert-butyl Ether	3.117	73	3145	1.041	ug/l	99
44) Trichloroethene	7.129	130	894	0.913	ug/l	77
64) Tetrachloroethene	9.268	164	382	0.443	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
Data File : VX045191.D
Acq On : 10 Mar 2025 13:20
Operator : JC/MD
Sample : Q1500-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW178I-20250305

Quant Time: Mar 11 03:54:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

