

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
 Data File : VX045196.D
 Acq On : 10 Mar 2025 15:16
 Operator : JC/MD
 Sample : Q1531-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D3-20250305

Quant Time: Mar 11 03:57:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	71783	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	142154	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	132743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	56805	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61586	53.937	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.880%			
35) Dibromofluoromethane	5.379	113	48788	51.326	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.660%			
50) Toluene-d8	8.647	98	177092	51.390	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.780%			
62) 4-Bromofluorobenzene	11.079	95	60678	53.137	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 106.280%			
Target Compounds						
16) Acetone	2.380	43	942	1.778	ug/l #	79
44) Trichloroethene	7.129	130	2379	2.461	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
Data File : VX045196.D
Acq On : 10 Mar 2025 15:16
Operator : JC/MD
Sample : Q1531-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D3-20250305

Quant Time: Mar 11 03:57:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

