

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
 Data File : VX045199.D
 Acq On : 10 Mar 2025 16:25
 Operator : JC/MD
 Sample : Q1531-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D2-20250306

Quant Time: Mar 11 03:58:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	76701	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	151163	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	137156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	56943	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63241	51.835	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.680%		
35) Dibromofluoromethane	5.379	113	51738	51.186	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.380%		
50) Toluene-d8	8.647	98	186578	50.916	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 101.840%		
62) 4-Bromofluorobenzene	11.079	95	63696	52.455	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 104.920%		
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.307	96	803	0.852	ug/l #	89
16) Acetone	2.380	43	677	1.196	ug/l #	86
44) Trichloroethene	7.123	130	402962	392.053	ug/l	100
64) Tetrachloroethene	9.275	164	572	0.651	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
Data File : VX045199.D
Acq On : 10 Mar 2025 16:25
Operator : JC/MD
Sample : Q1531-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D2-20250306

Quant Time: Mar 11 03:58:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

