

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
 Data File : VX045200.D
 Acq On : 10 Mar 2025 16:48
 Operator : JC/MD
 Sample : Q1531-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D3-20250306

Quant Time: Mar 11 03:58:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	69178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	138281	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51747	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58149	52.845	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.680%	
35) Dibromofluoromethane	5.385	113	48010	51.922	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.840%	
50) Toluene-d8	8.647	98	171889	51.277	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.560%	
62) 4-Bromofluorobenzene	11.079	95	57568	51.825	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.660%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	914	1.790	ug/l #	78
44) Trichloroethene	7.123	130	109525	116.487	ug/l	99
64) Tetrachloroethene	9.275	164	331	0.408	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
Data File : VX045200.D
Acq On : 10 Mar 2025 16:48
Operator : JC/MD
Sample : Q1531-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D3-20250306

Quant Time: Mar 11 03:58:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

