

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
 Data File : VX045201.D
 Acq On : 10 Mar 2025 17:12
 Operator : JC/MD
 Sample : Q1531-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D2-20250306

Quant Time: Mar 11 03:59:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	70057	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	136914	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50043	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59677	53.553	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.100%	
35) Dibromofluoromethane	5.385	113	48087	52.525	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.040%	
50) Toluene-d8	8.647	98	169351	51.024	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.040%	
62) 4-Bromofluorobenzene	11.079	95	56381	51.263	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.520%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.324	101	4252	5.222	ug/l	93
16) Acetone	2.379	43	783	1.514	ug/l #	84
44) Trichloroethene	7.122	130	345166	370.772	ug/l	99
64) Tetrachloroethene	9.268	164	1121	1.410	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
Data File : VX045201.D
Acq On : 10 Mar 2025 17:12
Operator : JC/MD
Sample : Q1531-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D2-20250306

Quant Time: Mar 11 03:59:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

