

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
 Data File : VX045202.D
 Acq On : 10 Mar 2025 17:35
 Operator : JC/MD
 Sample : Q1531-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D1-20250306

Quant Time: Mar 11 03:59:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	75048	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	150324	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	58912	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63579	53.260	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.520%	
35) Dibromofluoromethane	5.385	113	51874	51.607	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.220%	
50) Toluene-d8	8.647	98	188604	51.756	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.520%	
62) 4-Bromofluorobenzene	11.079	95	64031	53.025	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.060%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1164	2.102	ug/l #	87
44) Trichloroethene	7.123	130	32531	31.827	ug/l	99
64) Tetrachloroethene	9.275	164	1881	2.118	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
Data File : VX045202.D
Acq On : 10 Mar 2025 17:35
Operator : JC/MD
Sample : Q1531-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D1-20250306

Quant Time: Mar 11 03:59:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

