

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
 Data File : VX045203.D
 Acq On : 10 Mar 2025 17:58
 Operator : JC/MD
 Sample : Q1531-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE105D1-20250306

Quant Time: Mar 11 04:00:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	76394	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	149625	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57421	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64262	52.884	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.760%	
35) Dibromofluoromethane	5.385	113	51234	51.208	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.420%	
50) Toluene-d8	8.646	98	186697	51.472	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.940%	
62) 4-Bromofluorobenzene	11.079	95	64399	53.579	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.160%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1042	1.084	ug/l	84
9) 1,1,2-Trichlorotrifluo...	2.324	101	7256	8.172	ug/l	95
16) Acetone	2.385	43	953	1.690	ug/l	93
44) Trichloroethene	7.122	130	49357	48.514	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
Data File : VX045203.D
Acq On : 10 Mar 2025 17:58
Operator : JC/MD
Sample : Q1531-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE105D1-20250306

Quant Time: Mar 11 04:00:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

