

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
 Data File : VX045204.D
 Acq On : 10 Mar 2025 18:21
 Operator : JC/MD
 Sample : Q1531-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D1-20250307

Quant Time: Mar 11 04:00:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	71727	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	142598	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	131835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54697	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61843	54.204	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.400%		
35) Dibromofluoromethane	5.379	113	49316	51.720	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 103.440%		
50) Toluene-d8	8.647	98	178157	51.538	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.080%		
62) 4-Bromofluorobenzene	11.079	95	60912	53.176	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 106.360%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	2443	2.930	ug/l	88
12) 1,1-Dichloroethene	2.312	96	701	0.795	ug/l	89
16) Acetone	2.373	43	813	1.536	ug/l #	68
19) Methyl tert-butyl Ether	3.117	73	2614	0.884	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	2902	2.700	ug/l	94
44) Trichloroethene	7.123	130	190197	196.163	ug/l	99
64) Tetrachloroethene	9.275	164	9951	11.785	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
Data File : VX045204.D
Acq On : 10 Mar 2025 18:21
Operator : JC/MD
Sample : Q1531-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D1-20250307

Quant Time: Mar 11 04:00:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

