

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
 Data File : VX045206.D
 Acq On : 10 Mar 2025 19:07
 Operator : JC/MD
 Sample : Q1531-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20250306

Quant Time: Mar 11 04:01:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	72172	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	145463	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	133283	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54358	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61451	53.529	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.060%	
35) Dibromofluoromethane	5.379	113	48967	50.343	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.680%	
50) Toluene-d8	8.647	98	179471	50.895	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.800%	
62) 4-Bromofluorobenzene	11.079	95	61687	52.791	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.580%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.325	101	9288	11.073	ug/l	98
12) 1,1-Dichloroethene	2.312	96	1537	1.733	ug/l #	90
16) Acetone	2.380	43	960	1.802	ug/l	97
27) cis-1,2-Dichloroethene	4.489	96	1791	1.656	ug/l	90
44) Trichloroethene	7.123	130	596882	603.479	ug/l	98
64) Tetrachloroethene	9.275	164	6802	7.968	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX031025\
Data File : VX045206.D
Acq On : 10 Mar 2025 19:07
Operator : JC/MD
Sample : Q1531-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20250306

Quant Time: Mar 11 04:01:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

