

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
 Data File : VX007973.D
 Acq On : 11 Mar 2019 16:44
 Operator : JC/SP
 Sample : K1841-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20190307

Quant Time: Mar 12 08:57:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	260243	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	406025	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	382742	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	187170	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	140939	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
35) Dibromofluoromethane	5.50	113	132026	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
50) Toluene-d8	8.71	98	489271	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
62) 4-Bromofluorobenzene	11.13	95	178581	53.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.80%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
Data File : VX007973.D
Acq On : 11 Mar 2019 16:44
Operator : JC/SP
Sample : K1841-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20190307

Quant Time: Mar 12 08:57:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

