

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
 Data File : VX007975.D
 Acq On : 11 Mar 2019 17:37
 Operator : JC/SP
 Sample : K1793-01DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D1-20190305DL

Quant Time: Mar 12 09:01:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	237680	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	368658	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	345757	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	183609	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	128599	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
35) Dibromofluoromethane	5.50	113	119362	50.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.24%	
50) Toluene-d8	8.71	98	446724	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
62) 4-Bromofluorobenzene	11.13	95	168736	55.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.14%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.38	101	1574	0.750	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	2234	0.877	ug/l	87
29) Tetrahydrofuran	5.17	42	394	0.307	ug/l #	44
44) Trichloroethene	7.21	130	87977	32.124	ug/l	99
64) Tetrachloroethene	9.34	164	6377	2.226	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
Data File : VX007975.D
Acq On : 11 Mar 2019 17:37
Operator : JC/SP
Sample : K1793-01DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D1-20190305DL

Quant Time: Mar 12 09:01:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

