

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
 Data File : VX007977.D
 Acq On : 11 Mar 2019 18:30
 Operator : JC/SP
 Sample : K1793-03DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE131D3-20190305DL

Quant Time: Mar 12 09:04:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	237740	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	365570	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	339873	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	177267	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	127044	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
35) Dibromofluoromethane	5.50	113	117994	50.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.92%	
50) Toluene-d8	8.71	98	444131	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
62) 4-Bromofluorobenzene	11.13	95	163091	54.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.34%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.37	101	88384	42.120	ug/l	99
44) Trichloroethene	7.21	130	7497	2.761	ug/l	95
64) Tetrachloroethene	9.34	164	2511	0.892	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
Data File : VX007977.D
Acq On : 11 Mar 2019 18:30
Operator : JC/SP
Sample : K1793-03DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE131D3-20190305DL

Quant Time: Mar 12 09:04:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

