

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
 Data File : VX007979.D
 Acq On : 11 Mar 2019 19:23
 Operator : JC/SP
 Sample : K1841-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D2-20190308

Manual Integrations
 APPROVED

MMDadoda
 3/12/2019 3:23:48 PM

Quant Time: Mar 12 10:28:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	256722	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	400202	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	379763	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	188019	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	139082	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
35) Dibromofluoromethane	5.51	113	129937	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
50) Toluene-d8	8.71	98	480201	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
62) 4-Bromofluorobenzene	11.13	95	179357	54.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.82%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluoroet	2.39	101	23091	10.190	ug/l	100
12) 1,1-Dichloroethene	2.38	96	10921	5.000	ug/l	99
16) Acetone	2.45	43	4301	2.744	ug/l	98
24) 1,1-Dichloroethane	3.70	63	4211	0.930	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	10404	3.783	ug/l	89
30) Chloroform	5.21	83	7500	1.751	ug/l	95
38) Carbon Tetrachloride	5.77	117	5563	1.617	ug/l	94
44) Trichloroethene	7.23	130	9221686	3101.814	ug/l	98
49) 1,4-Dioxane	7.76	88	1202	17.487	ug/l	94
55) 1,1,2-Trichloroethane	9.21	97	5047m	1.893	ug/l	
64) Tetrachloroethene	9.34	164	6687	2.125	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
Data File : VX007979.D
Acq On : 11 Mar 2019 19:23
Operator : JC/SP
Sample : K1841-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D2-20190308

Manual Integrations
APPROVED

MMDadoda
3/12/2019 3:23:48 PM

Quant Time: Mar 12 10:28:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

