

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
 Data File : VX007980.D
 Acq On : 11 Mar 2019 19:50
 Operator : JC/SP
 Sample : K1841-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D1-20190308

Quant Time: Mar 12 09:10:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	255279	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	396303	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	379030	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	184121	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	140112	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
35) Dibromofluoromethane	5.51	113	129786	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
50) Toluene-d8	8.71	98	484242	52.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.02%	
62) 4-Bromofluorobenzene	11.13	95	177592	54.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	7986	3.544	ug/l	99
16) Acetone	2.45	43	4410	2.829	ug/l	94
21) trans-1,2-Dichloroethene	3.17	96	1055	0.441	ug/l #	76
27) cis-1,2-Dichloroethene	4.60	96	6222	2.275	ug/l	89
30) Chloroform	5.22	83	2277	0.535	ug/l	89
44) Trichloroethene	7.21	130	1519199	516.026	ug/l	98
64) Tetrachloroethene	9.34	164	4503	1.434	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
Data File : VX007980.D
Acq On : 11 Mar 2019 19:50
Operator : JC/SP
Sample : K1841-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE122D1-20190308

Quant Time: Mar 12 09:10:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

