

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
 Data File : VX007993.D
 Acq On : 12 Mar 2019 04:28
 Operator : JC/SP
 Sample : VX0311WBL03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0311WBL03

Quant Time: Mar 12 09:32:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	252220	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	389228	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	373573	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	183855	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	134548	50.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.14%	
35) Dibromofluoromethane	5.50	113	125032	50.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.44%	
50) Toluene-d8	8.71	98	478808	52.36	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.72%	
62) 4-Bromofluorobenzene	11.13	95	178271	55.61	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.22%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
Data File : VX007993.D
Acq On : 12 Mar 2019 04:28
Operator : JC/SP
Sample : VX0311WBL03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0311WBL03

Quant Time: Mar 12 09:32:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

