

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
 Data File : VX008003.D
 Acq On : 12 Mar 2019 08:54
 Operator : JC/SP
 Sample : K1793-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D1-20190306

Quant Time: Mar 12 13:45:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	249685	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387163	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	367423	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	179506	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	135969	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
35) Dibromofluoromethane	5.51	113	126369	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
50) Toluene-d8	8.71	98	472599	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
62) 4-Bromofluorobenzene	11.13	95	174205	54.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.26%	
Target Compounds						
16) Acetone	2.45	43	4718	3.095	ug/l	94
44) Trichloroethene	7.21	130	26519	9.220	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
Data File : VX008003.D
Acq On : 12 Mar 2019 08:54
Operator : JC/SP
Sample : K1793-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D1-20190306

Quant Time: Mar 12 13:45:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

