

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
 Data File : VX008005.D
 Acq On : 12 Mar 2019 09:47
 Operator : JC/SP
 Sample : K1793-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D3-20190306

Quant Time: Mar 12 13:45:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	246403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	387721	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	366873	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	177772	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	135050	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
35) Dibromofluoromethane	5.50	113	126252	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
50) Toluene-d8	8.71	98	473619	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
62) 4-Bromofluorobenzene	11.13	95	172528	54.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.06%	
Target Compounds						
16) Acetone	2.45	43	2836	1.885	ug/l #	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031119\
Data File : VX008005.D
Acq On : 12 Mar 2019 09:47
Operator : JC/SP
Sample : K1793-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D3-20190306

Quant Time: Mar 12 13:45:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

