

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
 Data File : VX015186.D
 Acq On : 11 Mar 2020 09:28
 Operator : JC/SP
 Sample : VX0311WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0311WBL01

Quant Time: Mar 12 06:27:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	327570	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	535603	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	501516	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	232108	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	207727	53.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.90%	
35) Dibromofluoromethane	5.48	113	172088	50.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.86%	
50) Toluene-d8	8.71	98	642228	49.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.94%	
62) 4-Bromofluorobenzene	11.13	95	219380	47.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
Data File : VX015186.D
Acq On : 11 Mar 2020 09:28
Operator : JC/SP
Sample : VX0311WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0311WBL01

Quant Time: Mar 12 06:27:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

