

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
 Data File : VX015189.D
 Acq On : 11 Mar 2020 10:53
 Operator : JC/SP
 Sample : L1842-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D1-20200305

Quant Time: Mar 12 07:24:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	332623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	535041	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	513300	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	249068	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	208638	52.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.74%	
35) Dibromofluoromethane	5.49	113	171687	50.86	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.72%	
50) Toluene-d8	8.71	98	650800	50.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.38%	
62) 4-Bromofluorobenzene	11.13	95	226828	49.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.06%	

Target Compounds						Qvalue
44) Trichloroethene	7.21	130	24517	6.00	ug/l	96
64) Tetrachloroethene	9.33	164	4326	1.02	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
Data File : VX015189.D
Acq On : 11 Mar 2020 10:53
Operator : JC/SP
Sample : L1842-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D1-20200305

Quant Time: Mar 12 07:24:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

