

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
 Data File : VX015190.D
 Acq On : 11 Mar 2020 11:15
 Operator : JC/SP
 Sample : L1860-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Mar 12 07:24:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	319599	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	525281	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	501329	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	248510	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	202942	53.52	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	107.04%	
35) Dibromofluoromethane	5.49	113	168316	50.79	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.58%	
50) Toluene-d8	8.71	98	635424	50.41	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	100.82%	
62) 4-Bromofluorobenzene	11.13	95	227990	50.71	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
12) 1,1-Dichloroethene	2.37	96	7145	2.30	ug/l	84
16) Acetone	2.43	43	6648	4.33	ug/l	89
24) 1,1-Dichloroethane	3.69	63	27831	4.65	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	84301	22.66	ug/l	89
32) 1,1,1-Trichloroethane	5.49	97	16355	3.38	ug/l	95
44) Trichloroethene	7.21	130	114219	28.47	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
Data File : VX015190.D
Acq On : 11 Mar 2020 11:15
Operator : JC/SP
Sample : L1860-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Mar 12 07:24:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

