

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
 Data File : VX015192.D
 Acq On : 11 Mar 2020 12:00
 Operator : JC/SP
 Sample : L1860-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Mar 12 07:25:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	300752	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	485779	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	483167	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	243697	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	174655	48.95	ug/l	0.00
Spiked Amount			Recovery	=	97.90%	
35) Dibromofluoromethane	5.49	113	152810	49.86	ug/l	0.00
Spiked Amount			Recovery	=	99.72%	
50) Toluene-d8	8.71	98	602238	51.66	ug/l	0.00
Spiked Amount			Recovery	=	103.32%	
62) 4-Bromofluorobenzene	11.13	95	222467	53.50	ug/l	0.00
Spiked Amount			Recovery	=	107.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	4279	2.96	ug/l	92
44) Trichloroethene	7.21	130	10269	2.77	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
Data File : VX015192.D
Acq On : 11 Mar 2020 12:00
Operator : JC/SP
Sample : L1860-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Mar 12 07:25:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

