

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
 Data File : VX015197.D
 Acq On : 11 Mar 2020 13:54
 Operator : JC/SP
 Sample : PB127431ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB127431ZHE#06

Quant Time: Mar 12 06:50:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	319992	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	544134	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	532193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	260959	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	204606	53.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.80%	
35) Dibromofluoromethane	5.49	113	171151	49.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.72%	
50) Toluene-d8	8.71	98	675363	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
62) 4-Bromofluorobenzene	11.13	95	243346	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	11619	7.56	ug/l	92
18) Methyl Acetate	2.76	43	4349	1.16	ug/l #	85
20) Methylene Chloride	2.85	84	9472	2.63	ug/l	95
43) Isopropyl Acetate	6.43	43	85410	10.40	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX031120\
Data File : VX015197.D
Acq On : 11 Mar 2020 13:54
Operator : JC/SP
Sample : PB127431ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB127431ZHE#06

Quant Time: Mar 12 06:50:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

